RESMED

Fitting Guide



- Ensure bottom of cushion opening aligns with the bottom of your nose then place the mask over your nose.
- Then pull headgear over your head so it sits midway between ears and eyes.



- Adjust the top strap first by pulling the strap through the buckle, one notch at a time, until the mask is comfortably over your nose.
- Do not overtighten.



- The back strap should sit just above the ears at the back of the head.
- Adjust both fasteners on the fabric back strap at the same time.
- Ensure back strap sits over your hair.



· Ensure mask is comfortably sealed.



- Adjust the nasal cushion until it sits comfortably on your nose.
- Connect the free end of the short tube assembly to the air tubing of your device, then turn it on.



- Check for leak and adjust if needed.
 If any air is leaking out then readjust the back strap before readjusting the top strap.
- Your mask is now ready to use.

Optional Soft Wraps



- Wrap each of the soft wraps around either side of the headgear, folding the top fastener over the lower one.
- The soft wraps should be positioned on your cheeks, with the fasteners on the outside.

Note: For additional information, refer to the Swift FX Nano / Swift FX Nano for Her Nasal Mask User Guide.

Swift[™] FX Nano



Distributed by: ResMed Corp 9001 Spectrum Center Boulevard San Diego CA 92123 USA. [ECIREP] ResMed (UK) Ltd 96 Milton Park Abingdon Oxfordshire 0X14 4RY UK. See www.resmed.com for other ResMed locations worldwide. Swift is a trademark of ResMed Ltd and is registered in U.S. Patent and Trademark Office. © 2013 ResMed Ltd. Specifications may change without notice. 1015934/1 2013-01

