RESMED

Mirage Micro[™] for Kids

NASAL MASK

Fitting Guide



• Turn the MicroFit dial to position 24, where the forehead support is fully extended.



• Disconnect one headgear clip, place the mask on the face and pull the headgear over the head.



- · Bring the lower straps below the ears.
- Connect the lower headgear clip ensuring it clicks into place.



- Unfasten the upper Velcro[™] tabs, pull the headgear straps evenly until comfortable, then reattach the Velcro.
- Do not overtighten the straps.



- Repeat Step 4 with the lower headgear straps.
- Connect one end of the air tubing to the swivel and the other end to the therapy device.
- Turn on the therapy device.



- While lying down, turn the MicroFit dial until the mask is evenly and comfortably positioned on the face.
- If any air leaks are felt, readjust the dial and/or headgear.

For details, refer to the Mirage Micro for Kids Nasal Mask User Guide.

ResMed Corp San Diego, CA, USA +1 858 836 5000 or 1 800 424 0737 (toll free). ResMed Ltd Bella Vista, NSW, Australia +61 (2) 8884 1000 or 1 800 658 189 (toll free). See www.resmed.com for other ResMed locations worldwide. Mirage Micro is a trademark of ResMed Ltd. Mirage is a trademark of ResMed Ltd and is registered in the U.S. Patent and Trademark Office. © 2010 ResMed Ltd. Specifications may change without notice. 1014018/1 10 04